



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/622,685	07/18/2003	Martin Frenz	02931-P0061A	3630

24126 7590 06/03/2005

ST. ONGE STEWARD JOHNSTON & REENS, LLC  
986 BEDFORD STREET  
STAMFORD, CT 06905-5619

EXAMINER

FARAH, AHMED M

ART UNIT PAPER NUMBER

3739

DATE MAILED: 06/03/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

e

<b>Office Action Summary</b>	<b>Application No.</b> 10/622,685	<b>Applicant(s)</b> FRENZ ET AL.	
	<b>Examiner</b> Ahmed M Farah	<b>Art Unit</b> 3739	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☐ Responsive to communication(s) filed on 11 March 2005.
- 2a) ☒ This action is **FINAL**.                      2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-44 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-44 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)  | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                                   | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)             |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)<br>Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____  |

## DETAILED ACTION

### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 2, 4-12, 20-33, and 40-44 are rejected under 35 U.S.C. 102(b) as being anticipated by Saadatmanesh et al. U. S. Patent No. 5,242,438.

Saadatmanesh et al. disclose an apparatus for applying light to a site to be treated on a wall of a vessel in the human body, the apparatus comprising: a light-feeding instrument designed to guide light generated by an external light source to said site and to radiate said light onto said site; the light-feeding instrument comprising an elongated light guide that can be inserted into said vessel and can be displaced therein in a longitudinal direction of said vessel, and light-deflecting means **178** designed to direct said light fed through said light guide in a substantially radial direction onto said site to be treated as presently claimed (see Figs. 1, 3, 4, and 6-12).

As to claims 5, 6 and 28, they teach that the angle of reflecting surface relative to the longitudinal axis of the catheter (i.e., the longitudinal axis of the optical fiber) is about 45°, see col. 8, lines 48-49.

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-44 are rejected under 35 U.S.C. 103(a) as being unpatentable over Saadatmanesh et al., in view of Kittrell et al. U. S. Patent No. 5,104,392.

Saadatmanesh et al., described above, do not teach a holding means for fixing the treatment apparatus within the blood vessel.

Kittrell et al. disclose an apparatus for applying light to a site to be treated on a wall of a vessel in the human body, the apparatus comprising: a light-feeding instrument designed to guide light generated by an external light source to said site and to radiate the treatment site and holding means (i.e., balloons) for holding the apparatus within the vessel during the treatment. The light-feeding instrument comprising an elongated light guide that can be inserted into said vessel and can be displaced therein in a longitudinal direction of said vessel; light-deflecting means designed to direct said light fed through said light guide in a substantially radial direction onto said site to be treated as presently claimed (see Figs. 1, 7E, 8, 9, 13A-13F, and 18-24).

Therefore, it would have been obvious to one skilled in the art at the time of the applicant's invention to modify the device of Saadatmanesh et al in view of Kittrell et al. and use positioning/holding balloons in order to fixate/position the treatment catheter within the treatment site.

### ***Conclusion***

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the date of this final action.


Any inquiry concerning this communication or earlier communications from the examiner should be directed to Ahmed M. Farah whose telephone number is (571) 272-4765. The examiner can normally be reached on Mon-Thur. 9:30 AM-7:30 PM, and 9:30 AM - 6:30 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda C.M DVorak can be reached on (571) 272-4768. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 3739

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Ahmed M Farah  
Primary Examiner  
Art Unit 3739



April 18, 2005.